

Total hoftealloplastik hos ældre – cementeret eller ikke-cementeret stem fiksation?

Anbefaling:

↑ Overvej anvendelse af cementeret stem fiksation til patienter over 70år, da risikoen for reoperation reduceres (+)

Det er arbejdsgruppens opfattelse, at litteraturen viser en forskel i risikoen for reoperation for kvinder helt ned til 60 år. Risikoen for reoperation bliver mere tydelig med stigende alder.

Udarbejdet af:

Dansk Selskab for Hofte- og Knæalloplastik (DSHK)

Forfattere:

Søren Overgaard (DHR)

Peter Horstmann (Gentofte)

Manuel Bieder (Næstved)

Martin Haubro (Odense)

Morten Bøgehøj (Aarhus)

Thomas Jakobsen (Aalborg)

Forventes godkendt forud for DOS kongressen 2023 efter høring på DOS-hjemmesiden.

Baggrund for valg af spørgsmål:

Total hoftealloplastik bliver i Danmark udført med cementeret eller ucementeret stem fiksation. En ucementeret stem fiksation er vist at være associeret til en øget risiko for tidlig reoperation grundet periprostetisk fraktur, mens cementeret stem fiksation i nogle studier har vist øget risiko for sen revision grundet aseptisk løsning.

Begge fiksations tekniker anvendes i dag, hvorfor DSHK har fundet det relevant at sammenligne fiksationsteknikerne for patienter over 60 år mht. mortalitet, risiko for reoperation, luksation, tromboemboliske komplikationer samt patient reporteret funktionsevne.

Denne retningslinje omhandler:

PICO spørgsmål:

Retningslinjen er udarbejdet med udgangspunkt i følgende PICO-spørgsmål:

Har patienter over 60 år med primær hofteledsartrose som indikation for indsættelse af en THA bedre effekt af cementeret stem fiksation end ucementeret stem fiksation med hensyn til reoperation samt funktionsevne?

Population: Alle patienter over 60 år i 10 år kohorter med primær hofteledsartrose som indikation for indsættelse af en total hoftealloplastik

Intervention: Cementeret stem fiksation

<u>Comparator:</u>	Ucementeret stem fiksation
<u>Outcome:</u>	Reoperation indenfor 2 år (Kritisk outcome) Reoperation indenfor 10 år (Kritisk outcome) Mortalitet indenfor 1 år (Kritisk outcome) Luksation indenfor 1 år (Sekundært outcome) Tromboemboliske komplikation indenfor 1 år (Sekundært outcome) Patient reporteret funktionsevne samt funktionsmåling efter 1 år (Sekundært outcome)

Anbefaling:

Følgende symboler, indikerer styrken af anbefalingerne:

↑↑ = Stærk anbefaling for

↑ = Svag/betinget anbefaling for

↓ = Svag/betinget anbefaling imod

↓↓ = Stærk anbefaling imod

✓ God praksis. Anvendes hvor der ikke findes evidens på området, men hvor arbejdsgruppen ønsker at fremhæve særlige aspekter af anerkendt klinisk praksis.

Følgende symboler angiver evidensniveau:

(+)(+)(+)(+) = Høj

(+)(+)(+) = Moderat

(+)(+) = Lav

(+) = Meget Lav

↑ Overvej anvendelse af cementeret stem fiksation til patienter over 70 år, da risikoen for reoperation reduceres (+)

Det er arbejdsgruppens opfattelse at litteraturen viser en forskel i risikoen for reoperation for kvinder helt ned til 60 år. Risikoen for reoperation bliver mere tydelig med stigende alder.

Litteratur:

Reoperation

Overordnet set tegner der sig et billede af lavere revisionsrate for cementerede THA og hybrid A. Dette gælder særligt for patienter over 70 år. Der synes at være en stigende risiko med stigende alder. Mænd har generelt en højere revisionsrate uafhængig af fiksationstype. De færreste studier opgør revisionsårsager, men periprostetiske frakturer synes at være en hyppigere tidlig komplikation for ucementeret teknik, mens aseptisk løsning fylder mest for cementeret teknik ved sen revision.

Mortalitet indenfor 1 år

De ikke justerede data viser generelt en højere mortalitet for cementeret stem fiksation som efter justering udlignes. Enkelte studier finder en let forøget mortalitet i de første postoperative dage ved cementeret stem fiksation, der er dog ikke konsensus omkring dette. Forskellen udligner sig ganske

hurtigt og efter et år ses ingen signifikant forskel på mortalitet. Mortaliteten synes ikke relateret til fiksationstypen, men i langt højre grad til alder, køn og comorbiditet.

Luksation indenfor 1 år

Hermansen et al. rapporterede, at cementeret fiksation beskytter mod luksation (OR 0,71 P = 0,001)) sammenlignet med ucementeret.

Tromboemboliske komplikation indenfor 1 år

Tromboemboli er en kendt komplikation efter operation med THA. I forhold til fiksationsmetoden konkluderes følgende:

Lindberg-Larsen et al. 2020, finder i deres opgørelse signifikant øget 30 dags postoperativ risiko for lungeemboli efter cementeret THA sammenlignet med ucementeret THA.

Ekman et al. 2019, finder sammenlignelig mortalitet i forhold til kardiovaskulære komplikationer mellem cementeret, hybrid og ucementeret THA.

McMinn et al. 2012, undersøgte med intraoperativt transøsophageal ekkokardiografi og viste forskel i embolisation af ekkogen materiale til højre hjertekammer og lungerne under operation med cementeret THA og resurfacing alloplastik. Operation med resurfacing alloplastik gav transient eller ingen embolisation, mens operation med cementeret THA gav betydelig embolisation i op til 20 minutter.

Patient reporteret funktionsevne samt funktionsmåling efter 1 år

Der er fundet 2 studier, der undersøger PROMS. Begge studier lader under meget små svarprocenter. Rolfsen et al. finder at der er bedre resultater for alle parametre i den ucementerede gruppe.

Jameson et al. finder at der ikke er nogen forskel mellem cementerede og ucementerede hofter, hverken i EQ5D eller HHS.

Da begge studier kigger på en meget lille andel af de producerede hofter i studieperioden, er det ikke muligt at konkludere noget ud fra disse studier.

Evidens:

Data er udelukkende indhentet fra register-studier, idet der ikke foreligger randomiserede studier af tilstrækkelig størrelse. Studierne er kendte tegnet ved store populationer, men typisk er grupperne uensartede i forhold til blandt andet alder og køn (case-mix), og nogle gange meget uens i størrelse. Der justeres for risikofaktorer, men hvilke varierer fra studie til studie. De statistiske metoder varierer en del, og nogle er baseret på kompleks modellering.

Reoperation

Evidensen er baseret på otte registerstudier. Der er i alle studier betydelig confounding og imprecision.

Mortalitet indenfor 1 år

Evidensen er baseret på seks registerstudier. Der er i alle studier betydelig confounding. Nogle studier sammenligner totalcementeret med ucementeret hvor andre sammenligner ucementeret med hybrid A. Evidens niveauet vurderes meget lavt (+)

Luksation indenfor 1 år

Luksationsrisikoen er vurderet ud fra et enkelt studie som opgør den sande luksationsrate i dansk setting.

Tromboemboliske komplikation indenfor 1 år

Evidensen er baseret på to registerstudier. Der er risiko for at resultaterne er påvirket af confounding.

Patient reporteret funktionsevne samt funktionsmåling efter 1 år

Studierne lider under meget lave svarprocenter, hvorfor det er vanskeligt at konkludere på. Evidensniveauet vurderes meget lavt (+)

I HØRING

Litteraturliste:

1. Babazadeh S, de Steiger RN, Holder C, van Bavel D. Cemented Polished Tapered Stems Have Lower Revision Rates Than Commonly Used Cementless Implant up to 17 Years of Follow-Up: An Analysis of 201,889 Total Hip Replacements From the Australian Orthopedic Association National Joint Replacement Registry. *J Arthroplasty*. 2022;37(1):110-8.
2. Bloemheuvel EM, Van Steenbergen LN, Swierstra BA. Comparable mortality but higher revision rate after uncemented compared with cemented total hip arthroplasties in patients 80 years and older: report of 43,053 cases of the Dutch Arthroplasty Register. *Acta Orthop*. 2022;93:151-7.
3. Dale H, Borsheim S, Kristensen TB, Fenstad AM, Gjertsen JE, Hallan G, et al. Perioperative, short-, and long-term mortality related to fixation in primary total hip arthroplasty: a study on 79,557 patients in the -Norwegian Arthroplasty Register. *Acta Orthop*. 2020;91(2):152-8.
4. Dale H, Borsheim S, Kristensen TB, Fenstad AM, Gjertsen JE, Hallan G, et al. Fixation, sex, and age: highest risk of revision for uncemented stems in elderly women - data from 66,995 primary total hip arthroplasties in the Norwegian Arthroplasty Register. *Acta Orthop*. 2020;91(1):33-41.
5. Ekman E, Palomaki A, Laaksonen I, Peltola M, Hakkinen U, Makela K. Early postoperative mortality similar between cemented and uncemented hip arthroplasty: a register study based on Finnish national data. *Acta Orthop*. 2019;90(1):6-10.
6. Garland A, Gordon M, Garellick G, Karrholm J, Skoldenberg O, Hailer NP. Risk of early mortality after cemented compared with cementless total hip arthroplasty: a nationwide matched cohort study. *Bone Joint J*. 2017;99-B(1):37-43.
7. Hermansen LL, Viberg B, Hansen L, Overgaard S. "True" Cumulative Incidence of and Risk Factors for Hip Dislocation within 2 Years After Primary Total Hip Arthroplasty Due to Osteoarthritis: A Nationwide Population-Based Study from the Danish Hip Arthroplasty Register. *J Bone Joint Surg Am*. 2021;103(4):295-302.
8. Jameson SS, Mason J, Baker PN, Gregg PJ, Deehan DJ, Reed MR. Implant Optimisation for Primary Hip Replacement in Patients over 60 Years with Osteoarthritis: A Cohort Study of Clinical Outcomes and Implant Costs Using Data from England and Wales. *PLoS One*. 2015;10(11):e0140309.
9. Kandala NB, Connock M, Pulikottil-Jacob R, Sutcliffe P, Crowther MJ, Grove A, et al. Setting benchmark revision rates for total hip replacement: analysis of registry evidence. *BMJ*. 2015;350:h756.
10. Kelly MP, Chan PH, Prentice HA, Paxton EW, Hinman AD, Khatod M. Cause-Specific Stem Revision Risk in Primary Total Hip Arthroplasty Using Cemented vs Cementless Femoral Stem Fixation in a US Cohort. *J Arthroplasty*. 2022;37(1):89-96 e1.
11. Lindberg-Larsen M, Petersen PB, Jorgensen CC, Overgaard S, Kehlet H, Lundbeck Foundation Center for Fast-track H, et al. Postoperative 30-day complications after cemented/hybrid versus cementless total hip arthroplasty in osteoarthritis patients > 70 years. *Acta Orthop*. 2020;91(3):286-92.
12. McMinn DJ, Snell KI, Daniel J, Treacy RB, Pynsent PB, Riley RD. Mortality and implant revision rates of hip arthroplasty in patients with osteoarthritis: registry based cohort study. *BMJ*. 2012;344:e3319.
13. Pedersen AB, Mailhac A, Garland A, Overgaard S, Furnes O, Lie SA, et al. Similar early mortality risk after cemented compared with cementless total hip arthroplasty for primary osteoarthritis: data from 188,606 surgeries in the Nordic Arthroplasty Register Association database. *Acta Orthop*. 2021;92(1):47-53.
14. Rolfson O, Donahue GS, Hallsten M, Garellick G, Karrholm J, Nemes S. Patient-reported outcomes in cemented and uncemented total hip replacements. *Hip Int*. 2016;26(5):451-7.
15. Stea S, Comfort T, Sedrakyan A, Havelin L, Marinelli M, Barber T, et al. Multinational comprehensive evaluation of the fixation method used in hip replacement: interaction with age in context. *J Bone Joint Surg Am*. 2014;96 Suppl 1(Suppl 1):42-51.
16. Tanzer M. Is cemented or cementless femoral stem fixation more durable in patients older than 75 years of age? A comparison of the best performing stems. *Clinical Orthopaedics and Related Research*. 2018.

Summary:

Abstract:

Short Clinical Guideline on total hip arthroplasty in the elderly, cemented or non-cemented stem fixation.

Background:

Total hip arthroplasties in Denmark is performed with either cemented or non-cemented stem fixation. The non-cemented stem has been showed to be associated to a higher risk of early reoperation due to periprosthetic fracture. The cemented stem has, in some studies, shown increased risk of late revision due to aseptic loosening.

Purpose / Aim of study:

Both cemented and non-cemented stems are used today, thus DSHK has found it relevant to compare these two fixation methods for patients above the age of 60 years, with regard to risk of reoperation, dislocation, risk of thromboembolic complication and patient reported outcome.

Materials and Methods:

The following PICO questions were investigated:

Does patients above 60 years of age with primary osteoarthritis operated with total hip arthroplasty have better effect of a cemented fixation than non-cemented stem fixation with regard to reoperation, mortality, dislocations within the 1st year, thromboembolic complications and functional outcome.

Findings/results

Reoperation:

Overall, there is a lower revision rate for cemented and hybrid A THA especially for patients above 75 years of age. Men tend to have a higher revision rate. Few studies report cause for revision, but periprosthetic fracture seems to be a common early complication for the non-cemented THA, whereas the aseptic loosening is more common for the cemented THA in the late revisions.

Mortality:

The non-adjusted numbers show a higher mortality for the cemented stem but after adjusting the numbers even out. Some studies find a higher mortality in the first days following surgery for cemented fixation but this evens out quite quick and after one year there is no difference between the two groups. Mortality seems not to be associated to fixation type, more to age, gender and comorbidities.

Dislocation within the 1st year

One study reported that cemented fixation prevents dislocation (OR 0,71 p=0,001) compared to non-cemented.

Thromboembolic complications

Thromboembolic complications is a well-known complication after THA.

One study showed significant increase in pulmonary embolism with in 30 days when comparing cemented to non-cemented. Another study showed no difference between the two groups.

Functional outcome:

Only two studies were found and due to the low number of included patients, the conclusion were very unclear.

Conclusions:

To consider the use of cemented fixation of the stem in patients above the age of 70 years since the risk of reoperation is reduced in cemented stem fixation.

The literature shows lower risk of revision in women down to the age of 60 years, the difference increases with rising age.

I HOPING

Bilag:

Bilag 1: Søgestrategi og søgestreng

Patient:

"arthroplasty, replacement, hip"[MeSH Terms] OR "Hip Prosthesis"[MeSH Terms] OR ("hip"[Text Word] AND ("arthroplast*"[Text Word] OR "prosthes*"[Text Word] OR "implant*"[Text Word] OR "replacement*"[Text Word])))

"osteoarthritis, hip"[MeSH Terms] OR "osteoarthritis*"[Text Word] OR "coxarthros*"[Text Word] OR "Arthroses"[Text Word] OR "arthrosis"[Text Word]

Intervention + Comparison:


"Bone Cements"[Pharmacological Action] OR "Bone Cements"[MeSH Terms] OR "Cementation"[MeSH Terms] OR "cement*"[Text Word] OR "uncement*"[Text Word] OR "noncement*"[Text Word]

Samlet søgestreng:

("arthroplasty, replacement, hip"[MeSH Terms] OR "Hip Prosthesis"[MeSH Terms] OR ("hip"[Text Word] AND ("arthroplast*"[Text Word] OR "prosthes*"[Text Word] OR "implant*"[Text Word] OR "replacement*"[Text Word]))) AND ("Bone Cements"[Pharmacological Action] OR "Bone Cements"[MeSH Terms] OR "Cementation"[MeSH Terms] OR "cement*"[Text Word] OR "uncement*"[Text Word] OR "noncement*"[Text Word]) AND ("osteoarthritis, hip"[MeSH Terms] OR "osteoarthritis*"[Text Word] OR "coxarthros*"[Text Word] OR "Arthroses"[Text Word] OR "arthrosis"[Text Word])

Søgning foretaget d. 9.marts 2023 i Pubmed og Embase.

Bilag 2: Flowskema over litteraturudvælgelse



Bilag 3: Vurdering af evidens

Quality assessment (GRADE)							
Design	Risk og bias	Inconsistency	Indirectness	Imprecision	Publication	Quality	
Outcome: Reoperation (Kritisk outcome)							
8 Registerstudier	Moderat til alvorlig risiko for bias	General enighed	Ja, der konkluderes ud fra fuld cementering i flere studier	Nej	Nej	Meget lav	Nedgraderet grundet risk of bias
Outcome: Mortalitet indenfor 1 år (Kritisk outcome)							
6 Registerstudier	Moderat til alvorlig risiko for bias	Ja, i de første postoperative døgn	Ja, der konkluderes ud fra fuld cementering i flere studier	Ja, brede confidensintervaller i hazrd ratio's	Nej	Meget lav	Nedgraderet grundet risk of bias
Outcome: Luksation indenfor 1 år (Sekundært outcome)							
1 Registerstudie	Moderat til alvorlig risiko for bias	Nej	Ja, der konkluderes ud fra fuld cementering. Hybrid A og B blandes.	Nej	Nej	Meget lav	Nedgraderet grundet risk of bias
Outcome: Tromboemboliske komplikation indenfor 1 år (Sekundært outcome)							
3 Registerstudier	Moderat til alvorlig risiko for bias	Ja	Der observeres på ultralydsfund i højre hjertekamre og lunge i et studie i stedet for klinisk emboli.	Nej	Nej	Meget lav	Nedgraderet grundet risk of bias
Outcome: Patient reporteret funktionsevne samt funktionsmåling efter 1 år (Sekundært outcome)							
2 Registerstudier	Moderat til alvorlig risiko for bias	Ja, forskellige resultater i de to studier	Nej	Nej	Nej	Meget lav	Nedgraderet grundet risk of bias

Bilag 4: Summary of Findings table

Reoperation

Authors	Results	Conclusions	Groups	n in groups	2-year results		10-year results				
							Male >70	Female >70			
1 Kandala et al.	1: Cemented/Hybrid lower revision than uncemented 2: Men higher revision-rates	1: 10-year revision-rates <5% for all types of THAs	CelMoP CelMoP CelCoC HyMoP CeCop All	125285 37874 34754 28471 12705 239089			2,22 [2,06-2,38] 3,25 [2,85-3,66] 3,96 [3,41-4,51] 2,42 [2,03-2,81] 1,88 [1,37-2,40] 2,54 [2,41-2,67]	2,93 [2,62-3,28] 4,31 [3,66-5,07] 4,39 [3,58-5,37] 3,18 [2,54-3,98] 2,10 [1,39-3,16] 3,25 [3,02-3,50]	2,67 [2,44-2,92] 3,37 [2,92-3,88] 3,76 [3,07-4,61] 2,63 [2,17-3,18] 1,68 [1,17-2,41] 2,79 [2,62-2,97]		
2 Kelly et al.	1: Hybrid has a higher risk of septic revision 2: Hybrid has a lower risk of PPF 3: Hybrid has a higher risk of aseptic loosening	cementless <75 in female cemented >75 in female	Groups	n in groups	2-year results		10-year results				
			Hybrid Uncemented (index) Female / Male <5 Female / Male 65-75 Female / Male >75	4539 84291			Statistics: COX regression 3.1 3.4 3.4 / 5.9 3.0 / 3.6 2.9 / 2.6	Septic revision "HR 1.5 index "HR 2.18 [3yr] HR 6.26 (>3yr) "HR 2.9 / # "HR 6.2 / # "HR 2.88 / #	Loosening "HR 2.93 / # "HR 6.23 / # "HR 2.88 / #	Instability "HR 0.41 index "#/# "#/# "#/# "#/#	PPF "HR 0.41 index "#/# "#/# "#/# "HR 0.27 / #
3 McMinn et al.	Results	Conclusions Slightly higher revision in uncemented THA	Groups Cemented (index) Uncemented	n in groups 154896 120017	2-year results		10-year results				
							Statistics: Royston-Parmar adjusted HR "HR 0.58 [0.54-0.62]				
4 Stea et al	1: hybrid lower risk of revision for 45-65/65-75/75 2: female lower risk of revision compared to men (HR 0.8, p<0.001) 3: Head size >32mm lower risk of revision (HR 0.63, p<0.001)	Conclusions Avoid uncemented >75yr	Groups cemented hybrid (index) uncemented	n in groups 52792 60502 126148	2-year results		10-year results				
							Statistics: Linear mixed models 45-64 yr #	65-74yr #	>75yr #		
5 Tanzer et al.	1: >75yr at 3 m.lower revision in 3 best cemented HR 3.31, p<0.006 2: at 13 yrs no difference in revision 3: Male higher risk of revision in both groups 4: early revision due to fractures	Conclusions Proper training in cementless technique	Groups cemented stem >75yr uncemented stem >75yr	n in groups 26627 4523	2-year results 3 months "HR 3.31 [1.39-7.90]	(index cemented stem)	Statistics: Kaplan Meier /COX 13 years 2.00 # HR 3.7	10-year results			
6 Babazadeh et al.	1: Fracture was most common cause for revision in cemented (47%) 2: Loosening most common cause for revision in uncemented (39%) 3: early revision due to fracture in cementless group 4: hybrid lower revision in young patients	Conclusions use cement for all ages with polished stem	Groups Polished stem (hybrid) Uncemented stem (3 most popular)	n in groups 101028 100861	2-year results 2 weeks "HR 0.48	(index hybrid)	Statistics: Kaplan Meier /COX 17 year [all] "HR 0.70 [0.5-0.99] 6.4 [6.0-6.8] 10.3 [6.4-13.4]	10-year results <55 HR 0.74			
7 Bloemheuvel et al.	1: Higher revision in uncemented THA compares to cemented 2: Male higher revision risk 3: Uncemented stem mostly revised due to fracture 4: cemented stem mostly revised due to dislocation	Conclusions 1: cement might be safer for patients >80 years	Groups cemented hybrid reverse hybrid Uncemented	n in groups 22025 3243 987 16376	1-year results 1.3 1.4 4.0 2.0		Statistics: KM/ competing risk /COX >80 years ONLY 2.2 HR 1.00 2.0 HR 1.00 6.2 "HR 2.9 [2.1-3.8] 2.8 "1.6 [1.4-1.8]	5-year results			
8 Dale et al.	1: men higher risk of revision (HR 1.6) 2: uncemented higher risk of revision 3: uncemented more PPF 4: uncemented less aseptic 5: the older the higher risk of PPF	Conclusions Uncemented stem should not be used in women >55	Groups Cemented Uncemented Reverse hybrid Hybrid All	n in groups 25678 16006 23312 1999	2-year results		Statistics: Kaplan Meier /COX infection index M/P: #/#	dislocation index M/P: HR 2.6 / HR 1.8	aseptic loosening index M/P: #/#	fractures index M/F: #/#	
											1.20% 0.70% 0.70% 0.40%

* significant resultat
ikke-significant resultat

Mortalitet

	Cemented				Hybrid A				Uncemented				Adjusted HR mortality for cementing
	n	30 days mortality	90 days mortality	1 year mortality	n	30 days mortality	90 days mortality	1 year mortality	n	30 days mortality	90 days mortality	1 year mortality	
McMinn et al.	155.996								120.017				HR 1.1; 95% CI 1.07 to 1.16, P<0.001
Pedersen et al.	108.572	HR 0.94; 95% CI 0.71-1.3	HR 0.97; 95% CI 0.79-1.2						80.034				
Bloemheuvel 2022					3.036	0.4%; 95% CI 0.3-0.5	0.7%; 95% CI 0.5-0.9	2.3%; 95% CI 2.0-2.6	15.215	0.4%; 95% CI 0.3-0.5	0.8%; 95% CI 0.7-0.9	2.6%; 95% CI 2.5-2.7	HR 1.0; 95% CI 0.9-1.0
Dale 2020 IV					2.955				21.553				No difference in 30days, 90 days, 10 years
Ekman 2019					11.802	HR 0.8; 95% CI 0.4-1.4	HR 0.7; 95% CI 0.5-1.1	HR 01.2; 95% CI 0.9-1.5	38.477				
Garland 2017*	146.818	HR 0.7; 95% CI 0.62-0.87	HR 0.5; 95% CI 0.40-0.50						15.410	HR 0.8; 95% CI 0.23-2.74	HR 0.2; 95% CI 0.07-0.56		

*Study comparing with matched controls from general population (HR 1.0)

Luksation

Author	Hybrid						Uncemented						
	n	in-hospital compl	30d readmis.	30d PE	30d mortality	30d dislocation	30d PPF	in-hospital compl	30d readmis.	30d PE	30d mortality	30d dislocation	30d PPF
Lindberg-Larsen et al.	8096	*7,7% (not multivariate)	#5,7%	*0,4% (uni/multi) HR=3,9	#0,2%	*1,2% (univariate)	*0,2% (univariate)	*5,3% (not multivariate)	#6,2%	*0,1% (uni/multi)	#0,3%	*1,8% (univariate)	*1,5% (univariate)

Tromboemboli

		cemented / hybrid stem fixation (n=3368)				uncemented stem fixation (n=4728)					
	n	30 days PE				30 days PE		p			
	8096	15 (0,4%), RR 3,9	(uni/multivariate)			4(0,1%)	(uni/multivariate)	<0,001			
Ekman et al, 2019		cemented fixation (n=23636)		hybrid fixation (n=11802)		uncemented fixation (n=38477)					
	n=73915	PE 0-10d		PE 0-10d		PE 0-10d					
		1 (0,00%)		0		2 (0,01%)					
		PE 90d		PE 90d		PE 90d					
		15 (0,06%)		4 (0,03%)		14 (0,04%)					
		PE 365d		PE 365d		PE 365d					
		30 (0,13%)		20 (0,05%)		7 (0,06%)					

PROM

Author	n	Cemented				Uncemented			
		EQ5D	EQ Vas	Pain VAS	Satisfaction VAS	EQ5D	EQ Vas	Pain VAS	Satisfaction VAS
Rolfson et al.	3,1+3,1k	*0,78	*76,6	*14,1	*14,6	*0,82	*79,4	*11,6	*12,1

* significant resultat

ikke-signifikant resultat

Tolkning: Ucementeret er bedre på alle parametre også når der justeres

I HØRING